
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000403
8:26 26/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:45
STERILIZING 5:08
136.4 C 231kPa 5:08
136.3 C 232kPa 5:38
136.4 C 231kPa 6:08
136.4 C 230kPa 6:38
136.0 C 230kPa 7:08
136.4 C 230kPa 7:38
136.0 C 229kPa 8:08
136.2 C 231kPa 8:38
136.2 C 231kPa 8:38
VENTING 8:38
AIR DRYING 9:33
CYCLE COMPLETE 9:35

Digital Signature #
03CBF1F68EE2F2A6
