
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001817
7:36 09/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:43
STERILIZING 3:44
136.1 C 220kPa 3:44
136.4 C 224kPa 4:14
136.4 C 224kPa 4:44
136.3 C 220kPa 5:14
136.4 C 220kPa 5:44
136.0 C 221kPa 6:14
136.4 C 224kPa 6:44
136.3 C 223kPa 7:14
136.3 C 223kPa 7:14
VENTING 7:14
AIR DRYING 8:05
CYCLE COMPLETE 22:56

Digital Signature #
897FACC9DEDD86A6
