

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001906  
8:19 29/06/2021

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:47  
PRESSURIZING 3:49  
STERILIZING 5:02  
136.4 C 222kPa 5:02  
136.3 C 223kPa 5:32  
136.1 C 222kPa 6:02  
136.2 C 222kPa 6:32  
136.4 C 221kPa 7:02  
136.4 C 224kPa 7:32  
136.3 C 220kPa 8:02  
136.1 C 221kPa 8:32  
136.1 C 221kPa 8:32  
VENTING 8:32  
AIR DRYING 9:26  
CYCLE COMPLETE 9:35  
-----

Digital Signature #  
52F42BD1EC2895C7  
-----