
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007767
7:22 20/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 3:00
STERILIZING 3:59
136.5 C 222kPa 3:59
136.4 C 222kPa 4:29
136.4 C 222kPa 4:59
136.4 C 221kPa 5:29
136.3 C 219kPa 5:59
135.9 C 217kPa 6:29
136.4 C 220kPa 6:59
136.4 C 222kPa 7:29
136.4 C 222kPa 7:29
VENTING 7:30
AIR DRYING 8:18
CYCLE COMPLETE 35:11

Digital Signature #
6A795FED125A39CB
