
Midmark M11 - v1.0.5
Total Cycles: 3781

Name: RITTER2
Sterilizer ID: V2313228
07/15/2024 11:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.5 F 0.0
2:00 139.1 F 1.7
4:00 147.8 F 4.5
6:00 168.4 F 8.1
8:00 189.6 F 12.3
10:00 214.2 F 5.8
12:00 226.1 F 10.4
14:00 231.2 F 8.1
16:00 240.0 F 12.4
18:00 239.4 F 10.9
20:00 251.7 F 17.0
22:00 262.5 F 23.4
24:00 272.7 F 30.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 30.0
1:00 272.1 F 30.3
2:00 272.7 F 30.3
3:00 272.8 F 30.2
4:00 272.7 F 30.1

Min 271.7 F 29.8
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 213.1 F 0.0
10:00 215.5 F 0.0
15:00 206.3 F 0.0
20:00 188.2 F 0.0
25:00 187.2 F 0.0
30:00 195.3 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:59
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:03:41
CYCLE COMPLETE

07/15/2024 12:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____