

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4247

Name: RITTER2  
Sterilizer ID: V2313228  
01/08/2025 11:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 142.1 F 0.0  
2:00 149.0 F 0.0  
4:00 150.6 F 0.2  
6:00 175.2 F 1.2  
8:00 193.4 F 5.0  
10:00 209.5 F 8.9  
12:00 223.8 F 13.4  
14:00 231.2 F 8.9  
16:00 240.0 F 11.7  
18:00 242.6 F 12.2  
20:00 241.4 F 11.3  
22:00 254.1 F 17.7  
24:00 265.2 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.2 F 29.1  
2:00 272.3 F 29.5  
3:00 271.7 F 28.6  
4:00 272.5 F 29.3

Min 271.5 F 28.6  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.4 F 0.0  
5:00 237.3 F 0.0  
10:00 227.8 F 0.0  
15:00 228.8 F 0.0  
20:00 226.0 F 0.0  
25:00 231.3 F 0.0  
30:00 230.4 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 25:50  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:04:16  
CYCLE COMPLETE

01/08/2025 12:30 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_