
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006095
10:15 10/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:26
STERILIZING 3:22
136.1 C 220kPa 3:22
136.4 C 225kPa 3:53
136.1 C 222kPa 4:23
136.4 C 225kPa 4:53
136.3 C 223kPa 5:23
136.4 C 225kPa 5:53
136.4 C 224kPa 6:23
136.4 C 224kPa 6:53
136.4 C 224kPa 6:53
VENTING 6:53
AIR DRYING 7:42
CYCLE COMPLETE 16:33

Digital Signature #
726D7CD61A8CA9EA
