
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000670
9:20 21/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:36
STERILIZING 4:39
136.0 C 220kPa 4:39
136.0 C 221kPa 5:09
136.3 C 224kPa 5:39
136.4 C 224kPa 6:09
136.2 C 220kPa 6:39
136.4 C 223kPa 7:09
136.1 C 221kPa 7:39
136.3 C 222kPa 8:09
136.3 C 222kPa 8:09
VENTING 8:10
AIR DRYING 9:03
CYCLE COMPLETE 9:07

Digital Signature #
DA2D37E105722BF2
