

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 006941  
12:37 13/02/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:13  
STERILIZING 4:08  
136.4 C 221kPa 4:08  
136.2 C 221kPa 4:39  
136.4 C 223kPa 5:09  
136.3 C 223kPa 5:39  
136.4 C 225kPa 6:09  
136.4 C 224kPa 6:39  
136.3 C 223kPa 7:09  
136.3 C 221kPa 7:39  
136.3 C 221kPa 7:39  
VENTING 7:39  
AIR DRYING 8:32  
CYCLE COMPLETE 9:06

-----  
Digital Signature #  
6FF081E157FF6033  
-----