
Midmark M11 - v1.0.5
Total Cycles: 3820

Name: RITTER2
Sterilizer ID: V2313228
07/29/2024 09:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.4 F 0.0
2:00 68.7 F 0.0
4:00 80.6 F 0.0
6:00 108.2 F 0.1
8:00 136.6 F 0.5
10:00 167.4 F 3.2
12:00 200.2 F 6.2
14:00 197.4 F 10.2
16:00 219.9 F 9.1
18:00 222.5 F 7.6
20:00 231.9 F 11.7
22:00 232.4 F 8.1
24:00 243.1 F 13.3
26:00 241.5 F 11.5
28:00 253.0 F 17.3
30:00 263.3 F 23.3
32:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.6 F 29.4
2:00 271.1 F 29.0
3:00 272.4 F 29.3
4:00 271.1 F 28.9

Min 271.0 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.7 F 0.0
5:00 237.6 F 0.0
10:00 179.5 F 0.0
15:00 175.7 F 0.0
20:00 171.4 F 0.0
25:00 169.5 F 0.0
30:00 168.4 F 0.0

Drying Complete

FILLING: 0:56
HEATING: 32:22
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:10:41
CYCLE COMPLETE

07/29/2024 10:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____