
Midmark M11 - v1.0.5
Total Cycles: 2224

Name: RITTER2
Sterilizer ID: V2313228
04/10/2023 01:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.3 F 0.0
2:00 139.6 F 1.9
4:00 155.3 F 5.1
6:00 194.4 F 9.6
8:00 216.2 F 6.2
10:00 225.0 F 8.7
12:00 236.6 F 11.5
14:00 238.8 F 11.2
16:00 237.3 F 9.5
18:00 251.0 F 16.2
20:00 262.7 F 23.0
22:00 272.6 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.3 F 29.5
2:00 271.8 F 29.6
3:00 272.2 F 29.3
4:00 272.0 F 29.7

Min 271.3 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.3 F 0.0
5:00 228.6 F 0.0
10:00 226.1 F 0.0
15:00 226.9 F 0.0
20:00 229.9 F 0.0
25:00 228.9 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:13
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:37
CYCLE COMPLETE

04/10/2023 02:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____