
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002295
11:54 19/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:02
STERILIZING 3:06
136.4 C 220kPa 3:06
136.3 C 225kPa 3:36
136.1 C 223kPa 4:06
136.4 C 224kPa 4:36
136.0 C 221kPa 5:06
136.4 C 225kPa 5:36
136.3 C 224kPa 6:06
136.0 C 221kPa 6:36
136.0 C 221kPa 6:36
VENTING 6:36
AIR DRYING 7:26
CYCLE COMPLETE 8:10

Digital Signature #
892099263F681F1F
