

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 003897  
8:47 28/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:39  
PRESSURIZING 2:41  
STERILIZING 3:40  
136.5 C 221kPa 3:40  
136.3 C 223kPa 4:10  
136.4 C 223kPa 4:40  
136.0 C 221kPa 5:10  
135.9 C 220kPa 5:40  
136.4 C 224kPa 6:10  
136.3 C 222kPa 6:40  
136.4 C 223kPa 7:10  
136.4 C 223kPa 7:10  
VENTING 7:10  
AIR DRYING 7:56  
CYCLE COMPLETE 8:03

-----  
Digital Signature #  
D318DDF7A4DFB459  
-----