

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
1.0uS / 0.6ppm  
CYCLE NUMBER 003769  
15:00 06/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:38  
STERILIZING 3:47  
136.5 C 224kPa 3:47  
136.4 C 226kPa 4:17  
136.1 C 223kPa 4:47  
136.4 C 223kPa 5:17  
136.4 C 225kPa 5:47  
136.3 C 223kPa 6:17  
136.4 C 222kPa 6:47  
136.4 C 225kPa 7:17  
136.4 C 225kPa 7:17  
VENTING 7:17  
AIR DRYING 8:55  
CYCLE COMPLETE 9:02

-----  
Digital Signature #  
BD9710373069C84B  
-----