

-----  
Midmark M11 - v1.0.5  
Total Cycles: 457

Name: RITTER2  
Sterilizer ID: V2313228  
10/18/2021 11:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.0 F 0.0  
2:00 156.7 F 2.0  
4:00 157.5 F 5.5  
6:00 182.4 F 10.2  
8:00 206.8 F 5.4  
10:00 218.4 F 10.2  
12:00 231.5 F 8.1  
14:00 241.1 F 12.7  
16:00 242.8 F 12.7  
18:00 255.9 F 19.4  
20:00 267.0 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 30.0  
1:00 271.8 F 30.0  
2:00 271.6 F 29.9  
3:00 271.3 F 30.0  
4:00 271.7 F 30.3

Min 271.2 F 29.9  
Max 272.5 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.3 F 0.0  
5:00 233.9 F 0.0  
10:00 234.7 F 0.0  
15:00 224.1 F 0.0  
20:00 222.6 F 0.0  
25:00 224.7 F 0.0  
30:00 226.5 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 21:26  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 00:59:33  
CYCLE COMPLETE

10/19/2021 06:12 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01