
Midmark M11 - v1.0.5
Total Cycles: 2053

Name: RITTER2
Sterilizer ID: V2313228
02/15/2023 08:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.6 F 0.0
2:00 81.3 F 0.0
4:00 102.6 F 0.0
6:00 147.1 F 0.0
8:00 161.0 F 0.0
10:00 184.5 F 2.0
12:00 201.7 F 4.9
14:00 214.3 F 8.2
16:00 228.4 F 11.9
18:00 228.8 F 7.4
20:00 240.0 F 12.4
22:00 238.4 F 10.3
24:00 237.1 F 8.7
26:00 250.0 F 15.5
28:00 261.5 F 22.2
30:00 272.1 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.4 F 29.2
2:00 272.1 F 29.4
3:00 272.1 F 28.9
4:00 272.6 F 29.6

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 205.0 F 0.0
10:00 219.4 F 0.0
15:00 224.6 F 0.0
20:00 232.0 F 0.0
25:00 221.7 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:29
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:08:32
CYCLE COMPLETE

02/15/2023 09:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____