
Midmark M11 - v1.0.0
Total Cycles: 3470

Name: RITTER 1
Sterilizer ID: V2178198
06/30/2023 09:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 215.3 F 0.0
2:00 204.3 F 2.2
4:00 216.6 F 6.5
6:00 231.9 F 11.9
8:00 229.4 F 7.7
10:00 243.3 F 13.3
12:00 240.2 F 10.7
14:00 236.0 F 8.6
16:00 250.7 F 15.5
18:00 262.2 F 22.2
20:00 272.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.8 F 29.0
2:00 271.8 F 28.3
3:00 271.5 F 28.6
4:00 272.8 F 29.0

Min 271.1 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 231.4 F 0.0
10:00 234.1 F 0.0
15:00 230.1 F 0.0
20:00 229.0 F 0.0
25:00 229.9 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 20:23
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:16
CYCLE COMPLETE

06/30/2023 10:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____