
Midmark M11 - v1.0.0
Total Cycles: 5302

Name: RITTER 1
Sterilizer ID: V2178198
01/30/2025 04:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.6 F 0.0
2:00 181.1 F 3.9
4:00 194.4 F 8.6
6:00 223.8 F 10.1
8:00 221.6 F 8.0
10:00 235.1 F 13.4
12:00 236.3 F 10.6
14:00 235.1 F 8.6
16:00 250.2 F 15.6
18:00 262.2 F 22.4
20:00 272.6 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.5 F 29.4
2:00 271.8 F 29.2
3:00 272.4 F 29.3
4:00 271.5 F 29.1

Min 271.4 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 230.4 F 0.0
10:00 252.3 F 0.0
15:00 240.4 F 0.0
20:00 218.5 F 0.0
25:00 211.1 F 0.0
30:00 207.4 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 20:24
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 00:59:32
CYCLE COMPLETE

01/30/2025 05:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____