
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000863
13:21 13/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:53
STERILIZING 4:03
136.4 C 233kPa 4:03
136.2 C 230kPa 4:33
136.4 C 231kPa 5:03
136.3 C 233kPa 5:33
136.4 C 232kPa 6:03
136.3 C 232kPa 6:33
136.1 C 230kPa 7:03
136.1 C 229kPa 7:33
136.1 C 229kPa 7:33
VENTING 7:33
AIR DRYING 8:24
CYCLE COMPLETE 23:24

Digital Signature #
0861656875B7E4EC
