

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5156

Name: RITTER 1  
Sterilizer ID: V2178198  
12/04/2024 12:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.1 F 0.0  
2:00 158.6 F 0.2  
4:00 174.0 F 0.7  
6:00 196.8 F 2.0  
8:00 217.0 F 6.0  
10:00 226.6 F 10.4  
12:00 230.3 F 7.5  
14:00 240.5 F 12.0  
16:00 238.8 F 10.0  
18:00 236.8 F 8.6  
20:00 250.9 F 15.6  
22:00 262.9 F 22.5  
24:00 272.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.2 F 28.8  
2:00 272.8 F 29.0  
3:00 272.4 F 28.7  
4:00 271.9 F 28.3

Min 271.3 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.0 F 0.0  
5:00 192.7 F 0.0  
10:00 192.9 F 0.0  
15:00 194.3 F 0.0  
20:00 197.4 F 0.0  
25:00 214.7 F 0.0  
30:00 219.4 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 24:22  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:02:53  
CYCLE COMPLETE

12/04/2024 01:22 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_