
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007647
9:38 10/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 1:57
STERILIZING 3:03
136.4 C 220kPa 3:03
136.4 C 224kPa 3:33
136.4 C 224kPa 4:03
136.2 C 222kPa 4:33
136.4 C 223kPa 5:03
136.2 C 219kPa 5:33
136.4 C 225kPa 6:03
136.4 C 223kPa 6:33
136.4 C 223kPa 6:33
VENTING 6:34
AIR DRYING 7:21
CYCLE COMPLETE 27:37

Digital Signature #
C9F268DAC9A0F67B
