
Midmark M11 - v1.0.0
Total Cycles: 1594

Name: RITTER 1
Sterilizer ID: V2178198
02/03/2022 07:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.1 F 0.0
2:00 75.6 F 0.0
4:00 88.2 F 0.1
6:00 117.3 F 0.1
8:00 140.9 F 2.3
10:00 169.1 F 5.1
12:00 193.6 F 8.5
14:00 218.6 F 12.2
16:00 220.4 F 4.8
18:00 232.4 F 8.8
20:00 241.6 F 13.0
22:00 237.0 F 9.6
24:00 239.2 F 9.4
26:00 246.8 F 13.3
28:00 257.9 F 19.5
30:00 268.5 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 271.9 F 28.4
2:00 272.3 F 28.9
3:00 271.2 F 28.4
4:00 272.0 F 28.5

Min 271.2 F 28.2
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.1 F 0.0
5:00 205.0 F 0.0
10:00 182.0 F 0.0
15:00 180.9 F 0.0
20:00 179.1 F 0.0
25:00 177.7 F 0.0
30:00 186.8 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 31:08
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:09:23
CYCLE COMPLETE

02/03/2022 09:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____