
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001170
12:43 18/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:09
PRESSURIZING 4:10
STERILIZING 5:34
136.0 C 220kPa 5:34
136.4 C 221kPa 6:04
136.3 C 222kPa 6:34
136.2 C 222kPa 7:04
136.4 C 224kPa 7:34
136.1 C 222kPa 8:04
136.1 C 221kPa 8:34
136.4 C 224kPa 9:04
136.4 C 224kPa 9:04
VENTING 9:04
AIR DRYING 10:00
CYCLE COMPLETE 25:00

Digital Signature #
452F1B58508303FD
