
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000122
9:46 04/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:29
PRESSURIZING 3:32
STERILIZING 4:38
136.0 C 220kPa 4:38
136.3 C 224kPa 5:08
135.9 C 221kPa 5:38
136.0 C 223kPa 6:08
136.3 C 223kPa 6:38
136.4 C 226kPa 7:08
136.3 C 224kPa 7:38
136.2 C 223kPa 8:08
136.2 C 223kPa 8:08
VENTING 8:08
AIR DRYING 9:03
CYCLE COMPLETE 24:03

Digital Signature #
8077F8F7BD6D7D2A
