
Midmark M11 - v1.0.0
Total Cycles: 2773

Name: RITTER 1
Sterilizer ID: V2178198
01/12/2023 10:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.6 F 0.0
2:00 202.1 F 0.7
4:00 220.2 F 6.4
6:00 230.9 F 11.9
8:00 231.8 F 8.1
10:00 242.1 F 11.6
12:00 243.1 F 12.4
14:00 242.6 F 11.5
16:00 256.4 F 18.6
18:00 268.3 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.9
2:00 271.7 F 28.1
3:00 272.1 F 28.8
4:00 272.3 F 28.7

Min 271.4 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.5 F 0.0
5:00 223.0 F 0.0
10:00 238.5 F 0.0
15:00 235.6 F 0.0
20:00 235.9 F 0.0
25:00 219.7 F 0.0
30:00 213.7 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

01/12/2023 11:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____