
Midmark M11 - v1.0.5
Total Cycles: 3298

Name: RITTER2
Sterilizer ID: V2313228
01/15/2024 11:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.7 F 0.0
2:00 140.6 F 0.8
4:00 149.4 F 3.0
6:00 175.4 F 6.5
8:00 194.6 F 10.7
10:00 224.0 F 5.8
12:00 232.7 F 9.5
14:00 235.9 F 10.2
16:00 238.3 F 11.3
18:00 237.7 F 9.6
20:00 250.7 F 15.9
22:00 262.0 F 22.4
24:00 272.4 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 271.3 F 29.2
2:00 271.3 F 29.2
3:00 271.4 F 29.1
4:00 271.6 F 29.2

Min 271.1 F 29.1
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.3 F 0.0
5:00 239.6 F 0.0
10:00 191.7 F 0.0
15:00 188.2 F 0.0
20:00 184.1 F 0.0
25:00 181.1 F 0.0
30:00 179.2 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 25:07
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:03:36
CYCLE COMPLETE

01/15/2024 12:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____