
Midmark M11 - v1.0.0
Total Cycles: 2233

Name: RITTER 1
Sterilizer ID: V2178198
08/18/2022 09:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.3 F 0.0
2:00 75.2 F 0.0
4:00 91.7 F 0.1
6:00 127.2 F 1.4
8:00 158.3 F 4.2
10:00 174.9 F 7.6
12:00 192.2 F 11.4
14:00 207.1 F 3.1
16:00 219.3 F 6.7
18:00 229.9 F 10.6
20:00 231.5 F 8.0
22:00 240.2 F 11.3
24:00 236.5 F 8.8
26:00 250.2 F 15.3
28:00 261.4 F 21.6
30:00 271.8 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.7 F 28.8
2:00 271.5 F 28.2
3:00 272.8 F 28.8
4:00 271.5 F 28.1

Min 271.4 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.7 F 0.0
5:00 210.5 F 0.0
10:00 202.5 F 0.0
15:00 205.1 F 0.0
20:00 197.6 F 0.0
25:00 201.3 F 0.0
30:00 188.2 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:35
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:08:29
CYCLE COMPLETE

08/18/2022 10:33 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____