

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002767  
9:16 01/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:01  
PRESSURIZING 3:03  
STERILIZING 4:08  
136.4 C 220kPa 4:08  
136.1 C 221kPa 4:38  
136.4 C 223kPa 5:08  
136.3 C 222kPa 5:38  
136.3 C 223kPa 6:08  
136.2 C 220kPa 6:38  
136.3 C 219kPa 7:08  
136.4 C 224kPa 7:38  
136.4 C 224kPa 7:38  
VENTING 7:38  
AIR DRYING 8:26  
CYCLE COMPLETE 8:31

-----  
Digital Signature #  
FEB08EE371445385  
-----