
Midmark M11 - v1.0.5
Total Cycles: 3802

Name: RITTER2
Sterilizer ID: V2313228
07/19/2024 01:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.0 F 0.0
2:00 93.4 F 0.0
4:00 109.5 F 0.1
6:00 145.7 F 1.5
8:00 177.1 F 4.2
10:00 192.3 F 7.5
12:00 208.9 F 10.9
14:00 220.3 F 6.5
16:00 227.9 F 8.4
18:00 236.1 F 12.4
20:00 233.9 F 9.0
22:00 242.1 F 11.8
24:00 243.5 F 12.5
26:00 254.2 F 18.1
28:00 264.1 F 24.0
30:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.6 F 29.8
2:00 272.2 F 29.4
3:00 271.1 F 28.9
4:00 271.8 F 29.7

Min 271.0 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 199.7 F 0.0
10:00 189.3 F 0.0
15:00 187.2 F 0.0
20:00 188.4 F 0.0
25:00 188.2 F 0.0
30:00 187.6 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 30:19
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:09:06
CYCLE COMPLETE

07/19/2024 02:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____