
Midmark M11 - v1.0.0
Total Cycles: 5206

Name: RITTER 1
Sterilizer ID: V2178198
12/19/2024 09:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.3 F 0.0
2:00 76.0 F 0.0
4:00 89.3 F 0.0
6:00 119.6 F 0.0
8:00 151.1 F 0.1
10:00 180.7 F 0.2
12:00 202.0 F 0.5
14:00 217.4 F 3.8
16:00 226.1 F 6.7
18:00 233.8 F 10.1
20:00 238.0 F 9.8
22:00 241.4 F 11.2
24:00 236.9 F 8.9
26:00 239.6 F 9.4
28:00 248.7 F 14.2
30:00 261.0 F 21.0
32:00 272.0 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.8
1:00 272.8 F 28.6
2:00 272.0 F 27.9
3:00 272.0 F 28.2
4:00 272.8 F 28.6

Min 271.5 F 27.7
Max 272.9 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 193.2 F 0.0
10:00 190.7 F 0.0
15:00 195.9 F 0.0
20:00 222.0 F 0.0
25:00 230.7 F 0.0
30:00 228.1 F 0.0

Drying Complete

FILLING: 0:40
HEATING: 32:36
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:10:39
CYCLE COMPLETE

12/19/2024 10:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____