

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1628

Name: RITTER 1  
Sterilizer ID: V2178198  
02/14/2022 11:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.9 F 0.0  
2:00 174.7 F 2.6  
4:00 184.1 F 6.4  
6:00 209.8 F 12.2  
8:00 222.8 F 6.5  
10:00 235.4 F 11.8  
12:00 236.0 F 9.1  
14:00 238.7 F 9.4  
16:00 248.8 F 14.5  
18:00 261.5 F 21.6  
20:00 272.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.2 F 28.6  
2:00 271.8 F 28.5  
3:00 272.1 F 28.4  
4:00 271.8 F 28.6  
  
Min 271.4 F 28.0  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 225.7 F 0.0  
10:00 232.4 F 0.0  
15:00 232.0 F 0.0  
20:00 228.9 F 0.0  
25:00 229.6 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 20:21  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:58:44  
CYCLE COMPLETE

02/14/2022 12:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_