
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001680
15:27 01/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:10
STERILIZING 4:16
136.4 C 221kPa 4:16
136.4 C 225kPa 4:46
136.3 C 225kPa 5:16
136.4 C 225kPa 5:46
136.0 C 223kPa 6:16
136.4 C 225kPa 6:46
136.4 C 225kPa 7:16
136.0 C 221kPa 7:46
136.0 C 221kPa 7:46
VENTING 7:46
AIR DRYING 8:40
CYCLE COMPLETE 8:55

Digital Signature #
1347FBBF07C83D216
