
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001373
9:58 01/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:19
PRESSURIZING 4:19
STERILIZING 5:42
136.0 C 220kPa 5:42
136.4 C 224kPa 6:12
136.2 C 223kPa 6:42
136.4 C 222kPa 7:12
136.3 C 224kPa 7:42
136.1 C 223kPa 8:12
136.3 C 224kPa 8:42
136.4 C 224kPa 9:12
136.4 C 224kPa 9:12
VENTING 9:12
AIR DRYING 10:11
CYCLE COMPLETE 21:48

Digital Signature #
F19B15620A67743D
