
Midmark M11 - v1.0.5
Total Cycles: 1896

Name: RITTER2
Sterilizer ID: V2313228
01/04/2023 04:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.0 F 0.0
2:00 172.2 F 3.6
4:00 188.9 F 7.9
6:00 209.7 F 13.2
8:00 225.0 F 8.0
10:00 238.2 F 13.6
12:00 240.4 F 11.7
14:00 239.6 F 10.7
16:00 253.7 F 17.7
18:00 265.5 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.4 F 29.7
2:00 271.6 F 28.9
3:00 272.1 F 29.9
4:00 272.1 F 29.4

Min 271.2 F 28.9
Max 272.4 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.7 F 0.0
5:00 230.1 F 0.0
10:00 229.1 F 0.0
15:00 227.1 F 0.0
20:00 218.7 F 0.0
25:00 217.2 F 0.0
30:00 217.6 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 19:44
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

01/04/2023 05:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____