
Midmark M11 - v1.0.0
Total Cycles: 5018

Name: RITTER 1
Sterilizer ID: V2178198
10/23/2024 03:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 87.2 F 0.0
2:00 90.6 F 0.0
4:00 104.8 F 0.1
6:00 137.8 F 0.3
8:00 170.8 F 1.0
10:00 187.1 F 4.0
12:00 200.5 F 7.1
14:00 213.9 F 10.3
16:00 222.0 F 9.7
18:00 228.9 F 8.5
20:00 238.1 F 13.1
22:00 238.3 F 10.2
24:00 236.9 F 8.7
26:00 247.8 F 14.1
28:00 258.6 F 20.1
30:00 268.6 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 271.7 F 28.8
2:00 272.3 F 28.9
3:00 271.6 F 28.3
4:00 271.6 F 28.8

Min 271.3 F 28.2
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.1 F 0.0
5:00 193.2 F 0.0
10:00 184.1 F 0.0
15:00 187.0 F 0.0
20:00 204.7 F 0.0
25:00 216.1 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:08
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:09:39
CYCLE COMPLETE

10/23/2024 04:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____