

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1989

Name: RITTER 1  
Sterilizer ID: V2178198  
06/01/2022 03:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.0 F 0.0  
2:00 159.4 F 3.4  
4:00 175.1 F 7.6  
6:00 198.1 F 12.4  
8:00 216.0 F 6.5  
10:00 229.7 F 11.4  
12:00 231.2 F 7.5  
14:00 242.7 F 13.3  
16:00 241.4 F 11.5  
18:00 254.7 F 18.1  
20:00 266.1 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.2 F 29.0  
2:00 272.1 F 29.4  
3:00 271.6 F 28.6  
4:00 272.5 F 29.3

Min 271.4 F 28.6  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.0 F 0.0  
5:00 204.6 F 0.0  
10:00 223.2 F 0.0  
15:00 237.2 F 0.0  
20:00 238.2 F 0.0  
25:00 226.1 F 0.0  
30:00 223.8 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 21:36  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:46  
CYCLE COMPLETE

06/01/2022 04:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_