
Midmark M11 - v1.0.0
Total Cycles: 5145

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2024 04:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.4 F 0.0
2:00 141.8 F 0.1
4:00 165.6 F 0.6
6:00 193.9 F 2.5
8:00 214.6 F 6.7
10:00 224.8 F 11.5
12:00 228.3 F 7.4
14:00 242.1 F 13.1
16:00 240.9 F 11.3
18:00 239.2 F 10.2
20:00 253.5 F 17.2
22:00 265.4 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.8 F 29.1
2:00 271.8 F 28.4
3:00 272.0 F 28.9
4:00 272.5 F 28.9

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 212.1 F 0.0
10:00 221.2 F 0.0
15:00 224.2 F 0.0
20:00 220.4 F 0.0
25:00 226.6 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 23:46
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:02:08
CYCLE COMPLETE

11/29/2024 05:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____