
Midmark M11 - v1.0.0
Total Cycles: 4777

Name: RITTER 1
Sterilizer ID: V2178198
07/31/2024 10:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.9 F 0.0
2:00 105.5 F 0.9
4:00 120.6 F 3.0
6:00 150.4 F 5.7
8:00 174.3 F 8.7
10:00 195.3 F 12.7
12:00 207.5 F 4.0
14:00 218.2 F 7.5
16:00 228.5 F 11.4
18:00 228.0 F 6.7
20:00 240.1 F 12.0
22:00 237.6 F 9.6
24:00 249.8 F 15.4
26:00 260.3 F 21.2
28:00 270.0 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 29.3
2:00 272.0 F 28.8
3:00 271.2 F 28.4
4:00 272.0 F 29.2

Min 271.2 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 199.6 F 0.0
10:00 196.0 F 0.0
15:00 198.5 F 0.0
20:00 218.9 F 0.0
25:00 233.5 F 0.0
30:00 233.2 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 29:03
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:07:36
CYCLE COMPLETE

07/31/2024 11:54 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____