
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001253
8:13 24/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:12
STERILIZING 4:23
136.4 C 222kPa 4:23
136.3 C 224kPa 4:53
136.3 C 223kPa 5:23
136.2 C 221kPa 5:53
136.0 C 220kPa 6:23
136.4 C 220kPa 6:53
136.4 C 219kPa 7:23
136.2 C 223kPa 7:53
136.2 C 223kPa 7:53
VENTING 7:53
AIR DRYING 8:44
CYCLE COMPLETE 8:46

Digital Signature #
0EE04FB925D8295C
