
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002145
12:49 27/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:05
STERILIZING 4:12
136.4 C 221kPa 4:12
136.4 C 225kPa 4:42
136.4 C 225kPa 5:12
136.3 C 221kPa 5:42
136.4 C 225kPa 6:12
136.3 C 222kPa 6:42
136.4 C 225kPa 7:12
136.1 C 220kPa 7:42
136.1 C 220kPa 7:42
VENTING 7:42
AIR DRYING 8:34
CYCLE COMPLETE 8:36

Digital Signature #
44597E636504B964
