
Midmark M11 - v1.0.5
Total Cycles: 30

Name: RITTER2
Sterilizer ID: V2313228
05/19/2021 10:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.5 F 0.0
2:00 148.3 F 2.4
4:00 173.0 F 5.9
6:00 190.8 F 10.5
8:00 212.5 F 4.8
10:00 224.0 F 9.5
12:00 232.3 F 8.8
14:00 240.7 F 12.2
16:00 240.4 F 10.9
18:00 253.2 F 17.4
20:00 264.4 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.2 F 29.2
2:00 272.5 F 29.7
3:00 271.5 F 29.0
4:00 272.5 F 29.5

Min 271.3 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.9 F 0.0
5:00 227.0 F 0.0
10:00 229.8 F 0.0
15:00 226.5 F 0.0
20:00 220.9 F 0.0
25:00 219.0 F 0.0
30:00 213.1 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:58
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:00:15
CYCLE COMPLETE

05/19/2021 11:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____