
Midmark M11 - v1.0.0
Total Cycles: 3980

Name: RITTER 1
Sterilizer ID: V2178198
11/03/2023 03:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.0 F 0.0
2:00 145.8 F 2.2
4:00 163.8 F 5.5
6:00 191.6 F 10.6
8:00 219.9 F 3.7
10:00 228.5 F 8.3
12:00 238.7 F 13.2
14:00 238.4 F 9.9
16:00 236.9 F 8.6
18:00 248.8 F 14.4
20:00 260.9 F 21.1
22:00 271.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 271.7 F 28.4
2:00 272.2 F 28.5
3:00 272.5 F 28.8
4:00 272.4 F 29.0

Min 271.3 F 28.4
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 227.3 F 0.0
10:00 235.5 F 0.0
15:00 229.9 F 0.0
20:00 225.4 F 0.0
25:00 214.8 F 0.0
30:00 213.2 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 23:10
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:01:52
CYCLE COMPLETE

11/03/2023 04:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____