
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001454
14:27 09/27/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:30
CYCLE ABORTED 6:12
WATER LEVEL LOW
NOT STERILE
END TIME 6:12

Digital Signature #
811E0C4FBEB277D8
