
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000358
13:09 20/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:08
PRESSURIZING 2:14
STERILIZING 3:16
136.0 C 222kPa 3:16
136.3 C 232kPa 3:47
136.3 C 232kPa 4:17
136.4 C 233kPa 4:47
136.4 C 232kPa 5:17
136.4 C 233kPa 5:47
136.4 C 232kPa 6:17
136.3 C 232kPa 6:47
136.3 C 232kPa 6:47
VENTING 6:47
AIR DRYING 7:39
CYCLE COMPLETE 8:19

Digital Signature #
F3784190E9374447
