
Midmark M11 - v1.0.0
Total Cycles: 5480

Name: RITTER 1
Sterilizer ID: V2178198
04/04/2025 02:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.2 F 0.0
2:00 132.0 F 1.7
4:00 157.2 F 4.6
6:00 175.8 F 7.9
8:00 201.8 F 11.7
10:00 214.8 F 3.5
12:00 222.7 F 7.5
14:00 232.8 F 11.7
16:00 231.2 F 7.2
18:00 243.0 F 12.4
20:00 240.1 F 10.3
22:00 252.0 F 16.2
24:00 262.7 F 22.3
26:00 272.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.3 F 28.7
2:00 271.5 F 28.5
3:00 272.0 F 28.7
4:00 271.3 F 28.4

Min 271.2 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 205.1 F 0.0
10:00 194.7 F 0.0
15:00 200.4 F 0.0
20:00 207.5 F 0.0
25:00 233.0 F 0.0
30:00 233.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 26:25
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:05:06
CYCLE COMPLETE

04/04/2025 03:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____