
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002088
14:45 23/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:37
STERILIZING 3:50
136.1 C 220kPa 3:50
136.1 C 223kPa 4:20
136.4 C 224kPa 4:50
136.4 C 226kPa 5:20
136.4 C 226kPa 5:50
136.4 C 226kPa 6:20
136.4 C 226kPa 6:50
136.3 C 225kPa 7:20
136.3 C 225kPa 7:20
VENTING 7:20
AIR DRYING 8:14
CYCLE COMPLETE 8:19

Digital Signature #
0B8941214789FE97
