
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000038
12:09 19/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:24
PRESSURIZING 1:29
STERILIZING 2:24
136.0 C 221kPa 2:24
136.0 C 221kPa 2:54
136.2 C 224kPa 3:24
136.2 C 224kPa 3:54
136.2 C 224kPa 4:24
136.4 C 220kPa 4:54
136.3 C 224kPa 5:24
136.4 C 222kPa 5:54
136.4 C 222kPa 5:54
VENTING 5:54
AIR DRYING 6:39
CYCLE COMPLETE 9:09

Digital Signature #
20EC3569ECEA7AC7
