
Midmark M11 - v1.0.5
Total Cycles: 1759

Name: RITTER2
Sterilizer ID: V2313228
11/22/2022 11:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.9 F 0.0
2:00 184.4 F 2.2
4:00 205.5 F 7.7
6:00 224.8 F 13.6
8:00 229.8 F 9.2
10:00 239.2 F 9.5
12:00 243.3 F 13.4
14:00 244.3 F 12.9
16:00 257.7 F 20.1
18:00 269.3 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.6 F 29.5
2:00 272.0 F 29.3
3:00 272.7 F 29.6
4:00 272.0 F 29.0

Min 271.8 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 214.1 F 0.0
10:00 218.4 F 0.0
15:00 223.2 F 0.0
20:00 223.8 F 0.0
25:00 222.5 F 0.0
30:00 217.0 F 0.0

Drying Complete

FILLING: 2:04
HEATING: 18:60
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:17
CYCLE COMPLETE

11/22/2022 12:06 PM
ACCEPTED: X
REJECTED:
Operator: U02