

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5203

Name: RITTER 1  
Sterilizer ID: V2178198  
12/18/2024 02:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.1 F 0.0  
2:00 138.8 F 0.0  
4:00 154.4 F 0.0  
6:00 182.3 F 0.3  
8:00 201.4 F 3.6  
10:00 220.4 F 7.3  
12:00 228.6 F 11.6  
14:00 229.5 F 7.3  
16:00 240.7 F 12.6  
18:00 239.7 F 10.5  
20:00 236.9 F 8.9  
22:00 251.0 F 15.6  
24:00 262.3 F 22.0  
26:00 272.3 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.3 F 28.7  
2:00 272.5 F 28.6  
3:00 271.9 F 28.2  
4:00 271.5 F 28.1

Min 271.3 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.5 F 0.0  
5:00 195.5 F 0.0  
10:00 191.1 F 0.0  
15:00 193.2 F 0.0  
20:00 220.3 F 0.0  
25:00 234.5 F 0.0  
30:00 230.5 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 26:34  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:05:06  
CYCLE COMPLETE

12/18/2024 03:49 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_