

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4369

Name: RITTER 1  
Sterilizer ID: V2178198  
03/08/2024 09:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.6 F 0.0  
2:00 75.2 F 0.2  
4:00 81.8 F 1.2  
6:00 110.0 F 3.1  
8:00 141.3 F 5.6  
10:00 164.2 F 8.5  
12:00 181.8 F 11.7  
14:00 201.4 F 1.9  
16:00 209.4 F 5.1  
18:00 218.7 F 8.2  
20:00 228.6 F 11.8  
22:00 228.9 F 7.0  
24:00 240.3 F 12.0  
26:00 238.2 F 9.6  
28:00 250.2 F 15.4  
30:00 260.4 F 21.1  
32:00 269.9 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.4 F 29.1  
2:00 272.1 F 28.6  
3:00 271.2 F 28.2  
4:00 272.4 F 29.2  
  
Min 271.1 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.2 F 0.0  
5:00 222.2 F 0.0  
10:00 230.0 F 0.0  
15:00 229.6 F 0.0  
20:00 227.9 F 0.0  
25:00 230.4 F 0.0  
30:00 228.4 F 0.0  
  
Drying Complete

FILLING: 0:51  
HEATING: 33:04  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:11:31  
CYCLE COMPLETE

03/08/2024 10:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_